








Determination of Cu-EDTA: linearity



Statistical data:	
number of x-levels	10
number of y-levels	10
sum of all x	616,00000000
sum of all y	11,17742540
quadratic sum x	81376,00000000
quadratic sum y	27,02271925
arithmetic mean x	61,60000000
slope m	0,01828982
intercept n	-0,00891042
coefficient of correlation R	0,99996543
determination coefficient R²	0,99993086
residual quadratic sum	0,00100453

- **Mode: CZE**
- **Electrolyte: Borate**
- **Capillary: 50 µm ID, 48 cm total, bubble cell**
- **Separation: -30 kV, 25°C**
- **Injection: hydrodynamic**
- **Detection: direct UV, 260 nm**