Determination of zinc and other cations in food supplements







- **Mode: CZE**
- **Electrolyte: aminopyridine**
- **Capillary: 50 µm ID, 64 cm total**
- **Separation: 30 kV, 25°C**
- **Injection: hydrodynamic, 50 mbar, 20 s**
- **Detection: indirect UV, 260 nm**

Operator: C. Amoussou, Chr. Öffler, 07-2021

© ICA 2021